



BDTİM



KANDİLLİ
RASATHANESİ VE
DEPREM ARAŞTIRMA
— ENSTİTÜSÜ —
1868

Kurulum Tarih : 14/05/2025

Görevli Personel : Mehmet Güler -Özkan Çok - Zafer Öğütcü

AYVL (AYVALIK - BALIKESİR) Deprem İstasyonu Kurulum Raporu

Serial-Nos=T6L51/3023

WO=13514

G=2392.34,2391.32,2387.26

VPC=0.962948,0.965911,0.963338

TYPE=6T

RESPONSE=CMG-6_30S_100HZ Vel

COILCONST=0.00438,0.00434,0.00423

CALVPC=0.58

CALRES=51000

İstasyon tipi: B-Band

Enlem: 39.2907N

Boylam: 26.6900E

Yükseklik : 17.0 m

Gprs : AVEA

Modem: Robustel

Hat No: 0505 176 15 14

IP: 176.30.211.241

Port:1567

Güneş paneli: 1 adet = 205 Watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

AYVALIK (AYVL) DEPREM İSTASYONU



Ayvalık Deprem İstasyonu yapım aşamaları







