



BDTİM



KANDİLLİ
RASATHANESİ VE
DEPREM ARAŞTIRMA
— ENSTİTÜSÜ —
1868

Tarih : 23/07/2025

Görevli Personel : Mehmet Güler - Burak Aksu – Rıdvan Özyılmaz

DIVR (SİVAS – DİVRİĞİ) İstasyonu Kurulum Raporu

Serial-Nos=T6Z70/C4561

WO=65662

G=2391.34,2389.08,2392.68

VPC=0.95989,0.960179,0.95986

TYPE=6T

RESPONSE=cmg-6_30s_100hz velocity

COILCONST=0.00450,0.00440,0.00434

CALVPC=0.58

CALRES=51000

İstasyon tipi: B-band

Enlem: 39.3562N

Boylam: 38.1237E

Yükseklik: 1257 m

1 Adet Guralp 6T sismometre

Güneş paneli: 1 adet = 205 Watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

1 adet Gprs modem



