



**BDTİM**



**KANDİLLİ**  
RASATHANESİ VE  
DEPREM ARAŞTIRMA  
— ENSTİTÜSÜ —  
1868

Tarih : 16/04/2025

Görevli Personel : Mehmet Güler-Murat Suvarıklı-Özkan Çok-Ferdi Olgun

**NMR2 ( Tekirdağ-Şarköy-Gaziköy-Uçmakdere) Kurulum Raporu**

Serial-Nos=T6M55/C3119

WO=13197

G=2394.14,2390.62,2393.1

VPC=0.964855,0.965273,0.966383

TYPE=6T

RESPONSE=CMG-6\_30S\_100HZ Vel

COILCONST=0.00449,0.00444,0.00419

CALVPC=0.58

CALRES=51000

Enlem: 40.7172

Boylam: 27.3123

Yükseklik: 10

Gprs : Turkcell

Modem: Cloudgate

Hat No: 0537 573 48 67

IP: 172.22.34.12

Port:1567

Güneş paneli: 2 adet : 75 watt x 50 watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

