



BDTİM



KANDİLLİ
RASATHANESİ VE
DEPREM ARAŞTIRMA
— ENSTİTÜSÜ —
1868

Tarih : 16/04/2025

Görevli Personel : Mehmet Güler-Murat Suvarıklı-Özkan Çok-Ferdi Olgun

NMR3 (Tekirdağ-Şarköy-Gaziköy-Uçmakdere) Kurulum Raporu

Serial-Nos=T6K90/C2922

WO=13292

G=2392.98,2397.12,2396.34

VPC=0.957854,0.954995,0.960946

TYPE=6T

RESPONSE=cmg-6_30s_100hz velocity

COILCONST=0.00452,0.00457,0.00431

CALVPC=0.96

CALRES=51000

Enlem: 40.7432

Boylam: 27.2944

Yükseklik: 235

Gprs : Turkcell

Modem: Cloudgate

Hat No: 0530 165 31 33

IP: 172.22.34.13

Port:1567

Güneş paneli: 2 adet : 75 watt x 50 watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

