



BDTİM



KANDİLLİ
RASATHANESİ VE
DEPREM ARAŞTIRMA
— ENSTİTÜSÜ —
1868

Tarih : 16/04/2025

Görevli Personel : Mehmet Güler-Murat Suvarıklı-Özkan Çok-Ferdi Olgun

NMR4 (Tekirdağ-Şarköy-Gaziköy-Uçmakdere) Kurulum Raporu

Serial-Nos=T6M41/C3118

WO=13197

G=2386.78,2392.84,2386.16

VPC=0.959122,0.962128,0.962396

TYPE=6T

RESPONSE=CMG-6_30S_100HZ Vel

COILCONST=0.00438,0.00447,0.00433

CALVPC=0.58

CALRES=51000

Enlem: 40.7880

Boylam: 27.3006

Yükseklik: 925

Gprs : Avea

Modem: Robustel

Hat No: 0553 042 87 92

IP: 176.30.204.219

Port:1567

Güneş paneli: 2 adet : 75 watt x 50 watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

