



BDTİM



KANDİLLİ
RASATHANESİ VE
DEPREM ARAŞTIRMA
— ENSTİTÜSÜ —
1868

Tarih : 16/04/2025

Görevli Personel : Mehmet Güler-Murat Suvarıklı-Özkan Çok-Ferdi Olgun

NMR5 (Tekirdağ-Şarköy-Gaziköy-Uçmakdere) Kurulum Raporu

Serial-Nos=T6K86/C2921

WO=13292

G=2381.96,2398.96,2399.24

VPC=0.956529,0.960296,0.963903

TYPE=6T

RESPONSE=cmg-6_30s_100hz velocity

COILCONST=0.00453,0.0046,0.00452

CALVPC=0.96

CALRES=51000

Enlem: 40.7954

Boylam: 27.3660

Yükseklik: 97

Gprs : Turkcell

Modem: Cloudgate

Hat No: 0533 042 48 23

IP: 172.22.34.1

Port:1567

Güneş paneli: 2 adet : 75 watt x 50 watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

