



BDTİM



KANDİLLİ
RASATHANESİ VE
DEPREM ARAŞTIRMA
— ENSTİTÜSÜ —
1868

Tarih : 16/04/2025

Görevli Personel : Mehmet Güler-Murat Suvarıklı-Özkan Çok-Ferdi Olgun

NMR6 (Tekirdağ-Şarköy-Gaziköy-Uçmakdere) Kurulum Raporu

Serial-Nos=T6K91/C2925

WO=13292

G=2391.94,2396.16,2389.88

VPC=0.965719,0.96658,0.964997

TYPE=6T

RESPONSE=cmg-6_30s_100hz velocity

COILCONST=0.00458,0.00459,0.0043

CALVPC=0.96

CALRES=51000

Enlem: 40.7040

Boylam: 27.2653

Yükseklik: 333

Gprs : Turkcell

Modem: UR75

Hat No: 0537 573 48 17

IP: 172.22.34.16

Port:1567

Güneş paneli: 2 adet : 75 watt x 50 watt

Güneş Paneli Regülatörü:20A

Akü: 1 adet 100 Amper

